

*Certificate
of
Accuracy*



PYRAMID GRANITE & METALS, INC.

Precision Products Since 1944

660 Superior St. • Escondido, CA • 92029 • Ph. (760) 745-6309 • Fax (760) 743-2803

www.pyramidgranite.com • 1-800-874-5441 • sales@pyramidgranite.com

ITEM: Granite Base Per Print CDI-6066 Rev. D

SERIAL NO.:

GRADE: AA

ACCURACY IN MICROINCHES: 50

REPEAT MEASUREMENT IN MICROINCHES:* 35

TOTAL FLATNESS TOLERANCE FOR WORKING SURFACE AREA
MEETS OR EXCEEDS REQUIREMENTS OF FED. SPEC. GGG-P-463c:**

This item was inspected under controlled conditions with electronic and optical gaging equipment calibrated and traceable to the National Institute of Standards and Technology (NIST) through Reference No. _____ applying to Master Gage Block Ser. No. RS-1.

Dated

Inspector

*Repeatability of height-gage measurements anywhere on surface area, Full Indicator Reading.

**Surface area includes the entire surface of the plate, except for the outer edge of the plate defined as 1" border for plates smaller than 48" diagonal, and 1.5" border for larger plates.