

OTHER MARK-10 PRODUCTS

A full line of force and torque measurement products and related items, including force gauges, torque gauges, force and torque test stands, sensors, grips, software, and more.



MARK-10

Force and torque measurement engineered better

Mark-10 Corporation

11 Dixon Avenue
Copiague, NY 11726
888-MARK-TEN
Tel: 631-842-9200
Fax: 631-842-9201
Internet: www.mark-10.com
E-mail: info@mark-10.com

32-1123
1122

Series TT01

CAP TORQUE TESTERS

Quick Start Guide

MARK-10

Thank you!



Thank you for purchasing a Mark-10 Series TT01 Cap Torque Tester. We are confident that you will get many years of service from this product.

Please read through this Quick Start Guide and the complete User's Guide before using the tester.


IMPORTANT SAFETY CONSIDERATIONS

Force and torque measuring instruments are sensitive and must be treated with care. Please review the safety information provided in the User's Guide prior to operating the gauge.

LIST OF INCLUDED ITEMS:

Qty.	Part No.	Description
1	MTT01-12 / MTT01-25 / MTT01-50 / MTT01-100	Series TT01 digital cap torque tester
1	AC1030 / AC1031 / AC1032 / AC1035	AC adapter/charger w/US, EU, UK, or AUS prong
1	CT004	Sample gripping posts, 4
1	AC1118	Rechargeable NiMH battery (inside the instrument)
1	-	Certificate of calibration
1	AC1111	USB cable, type micro B to A
1	CT001	Flat jaws set, 2 (optional)
1	CT002	Carrying case (optional)
1	CT003	Adjustable jaws set, 2 (optional)
1	CT005	Set of 4 posts, 2.5" (optional)
1	CT006	Set of 4 posts, 4" (optional)
1	CT007	Kit of 4 posts of each length (optional)

POWER

To power on the tester, press the  button. Press and hold to power off. The tester may be powered either by the internal rechargeable battery or an AC adapter/charger. The battery will last approximately 7 hours with the backlight turned on, and 24 hours with the backlight turned off before recharging is necessary.

LITERATURE & SOFTWARE DOWNLOADS

Download at: www.mark-10.com/downloads

- User's guides** (PDF)
Provided for all Mark-10 instruments. Locate the user's guide for this product and read it thoroughly before use.
- USB driver**
For Mark-10 instruments and the model RSU100 RS-232 to USB converter. Required for communication with a PC. Follow the installation instructions provided in the user's guide.
- MESUR® family of software** – see below.

INSTALLING MESUR® SOFTWARE

Mark-10 offers three data collection applications:

MESUR® Lite

Free basic data acquisition application, for data tabulation and exporting to Excel. Compatible with gauges only, not test stands.

MESUR®gauge / MESUR®gauge Plus

Data tabulation, graphing, analysis, and reporting, with user-configurable start and stop triggers and other tools. Compatible with gauges and test stands for load vs. time or load vs. travel. MESUR®gauge Plus adds test stand motion control coordination. Both applications can be evaluated for 90 days.

PC requirements

Compatible with PCs running Microsoft Windows 7 or later. Minimum monitor resolution is 1108 x 758. A USB or RS-232C serial port is required. If USB communication is required, install the Mark-10 USB driver.